

BOARD OF  
BUILDING AND SAFETY  
COMMISSIONERS

MARSHAL L. BROWN  
PRESIDENT

VAN AMBATILOS  
VICE-PRESIDENT

VICTOR H. CUEVAS

HELENA JUBANY

ELENORE A. WILLIAMS

CITY OF LOS ANGELES

CALIFORNIA



ANTONIO R. VILLARAIGOSA  
MAYOR

DEPARTMENT OF  
BUILDING AND SAFETY  
201 NORTH FIGUEROA STREET  
LOS ANGELES, CA 90012

ROBERT R. "BUD" OVROM  
GENERAL MANAGER

RAYMOND S. CHAN, C.E., S.E.  
EXECUTIVE OFFICER

JAMES CHANDLER  
VISTA RESEARCH GROUP, LLC  
1554 TOWNSHIP ROAD 805  
ASHLAND, OH 44805-9202

Telephone: (419) 281-3927

Models:

SINGLE OPERATORY: 1000 | CENTRAL SYSTEM: 1000-(23, 34, 46 & 47)-C.

File Number: M-040014

Effective Date: 08/03/2011

Expiration Date: 07/01/2012

Manufacturer: SAME

Item: DENTAL WATERLINE FILTER

**PLEASE REFER TO THE APPLICATION NUMBER ON ALL CORRESPONDENCE, PHONE CALLS & SAMPLES**

The item processed under the above file number has been tested, examined and found to comply with applicable Los Angeles Plumbing Code (LAPC) and/or Los Angeles Mechanical Code (LAMC) and is hereby approved unless revoked for cause.

This letter or the letterhead copy may not be reproduced for use in labeling or advertising, but copies of this letter may be used within your organization and shown as evidence of approval.

CONDITIONS OF APPROVAL:

1. An approved spring-loaded double check valve shall be installed on each water inlet to the filter per Section 94.603 of LAPC, 2011 Edition.
2. The drain line shall be routed without dips and sags into an indirect waste receptor 1" or two pipe size, whichever is greater, above its flood level per Section 94.801 of LAPC, 2011 Edition.
3. This product shall be installed, operated, and maintained in accordance with instructions from the manufacturer and LAPC, 2011 Edition.

Each item shall have the name or trademark of manufacturer and model number where it will be visible for inspection.

In order to renew this approval, an application and reexamination fee must be submitted on or before the expiration date shown above. Arrangement shall be made to remove any test samples remaining in the Laboratory. Samples remaining after 21 days from the date of this notice will be shipped collect by the most convenient carrier or disposed of in accordance with applicable regulation.

Thomas Liu  
Mechanical Testing Laboratory, Test Engineer

Plumbing/Mechanical Inspection  
Inspection Bureau